

Intrada Amino Acid

150 x 3 mm



Application

LC-TOFMS: Urinary compounds



尿代謝関連化合物のLC-TOFMS分析

Metabolomics Core Facility, McGill University
GB 2017/06/16

Argininosuccinic acid [M+H]⁺: 291.1299



Methylnicotinamide [M+H]⁺: 137.0709



Intrada Amino Acid, 150x3mm

A) ACN /HCOOH = 100 / 0.3, B) 100mM HCOONH₄ in 20% ACN / 80% water
0-27%B (0-3min), 27-100%B (3-19.5min), 100%B (19.5-25min), 0%B (25-32min)
0.6 mL/min, 10°C, 5μL (ca. 1mM in water)

Agilent 1290 Infinity UHPLC - 6540 UHD Accurate-Mass Q-TOF mass spectrometer (ESI⁺)

Courtesy of Dr.Gaelle Bridon and Dr.Daina Avizonis, McGill University, CANADA