


Presto FF-C18

150 x 4.6 mm

HPLC COLUMN APPLICATION

Hydroxypropyl methylcellulose


ヒドロキシプロピルメチルセルロース



hydroxypropyl methylcellulose
(40 ~ 50 cps, 2% aq.Soln.)

50-85%B (0-15min)
10 μ L (5mg/mL, water / AcOH = 100 / 1)

0-100%B (0-8min), 100%B (8-15min)
5 μ L (5mg/mL, water / AcOH = 100 / 1)



Presto FF-C18, 150 x 4.6 mm
A: water / acetic acid = 100 / 0.5
B: acetonitrile / water / acetic acid = 90 / 10 / 0.5
0.4mL/min (18MPa), 37°C
ELSD (spray chamber 30°C, drift tube 60°C)