


Unison UK-Amino

150 x 4.6 mm

HPLC COLUMN APPLICATION

Changes in gluconolactone over time

グルコノラクトンの経時変化



Unison UK-Amino, 150 x 4.6 mm

A: acetonitrile

B: 100mM ammonium acetate

0 -100%B (0 -15 min), 1.0 mL/min (6 MPa) , 37°C

ELSD (spray chamber 70°C, drift chamber 100°C)

10 µL (5.0mg/mL, 100mM ammonium acetate)